


CERTIFIED COPY (CPLR RULE 4518)

0.10% SIMULATOR SOLUTION RECORD

Lot Number 11190, dated October 18, 2011

I, Jennifer F Limoges, Supervisor of Forensic Services, Forensic Investigation Center, do hereby certify and authenticate, as provided by Rule 4518, subdivision c, Civil Practice Law and Rules, that the document annexed hereto is a true and accurate representation of an electronic record of the New York State Police Forensic Investigation Center which has been delegated to my possession, custody and control by the Superintendent of the New York State Police. This record was made in the regular course of business of the New York State Police Forensic Investigation Center, it is the regular course of the Forensic Investigation Center business to make such record, and such record was made at the time of each event recorded in it or within a reasonable time thereafter.



Digitally signed under
ESRA by Jennifer F Limoges
on
2011.10.24 13:45:39 -04'00'

Jennifer F Limoges

Supervisor of Forensic Services
Forensic Investigation Center



JOSEPH A. D'AMICO
SUPERINTENDENT

**NEW YORK STATE POLICE
FORENSIC INVESTIGATION CENTER
BUILDING 30
1220 WASHINGTON AVE.
ALBANY, NY 12226-3000**

0.10% SIMULATOR SOLUTION RECORD

Report Date: October 18, 2011
Manufacturer: Guth Laboratories, Inc.
Lot Number: 11190
Expiration Date: 3/28/2012
Analysis Results: 122 mg Ethyl Alcohol / 100 mL of solution

I, Carrie A Kirkton, tested simulator solution lot number 11190 by headspace gas chromatography and have determined that it contains the appropriate concentration of ethyl alcohol. This solution is hereby approved for use.

When this simulator solution is used with a properly operating breath testing instrument, it will provide a value of 0.10% within acceptable limits.


I, Carrie A Kirkton, made the entries contained in this record at the time of each event recorded, or within a reasonable time thereafter.

Performed by:

 Digitally signed under ESRA
by Carrie A Kirkton on
2011.10.18 13:50:17 -04'00'

Carrie A Kirkton
Forensic Scientist III
Breath Analysis Technical Supervisor

Reviewed by:

 Digitally signed under ESRA
by Jennifer F Limoges on
2011.10.24 13:44:58 -04'00'

Jennifer F Limoges
Supervisor of Forensic Services
Breath Analysis Technical Supervisor