


CERTIFIED COPY (CPLR RULE 4518)

0.10% SIMULATOR SOLUTION RECORD

Lot Number 15210, dated October 15, 2015

I, Jennifer F Limoges, Associate Director of Forensic Science, Forensic Investigation Center, do hereby certify and authenticate, as provided by Rule 4518, subdivisions a and c, Civil Practice Law and Rules, that the document annexed hereto is a true and accurate representation of an electronic record of the New York State Police Forensic Investigation Center which has been delegated to my possession, custody and control by the Superintendent of the New York State Police. This record, once created, is stored in an electronic format that cannot, thereafter, be altered or modified. This record was made in the regular course of business of the New York State Police Forensic Investigation Center, it is the regular course of the Forensic Investigation Center business to make such record, and such record was made at the time of each event recorded in it or within a reasonable time thereafter.

 Digitally signed under
ESRA by Jennifer F Limoges
on
2015.10.16 14:35:56 -04'00'

Jennifer F Limoges

Associate Director of Forensic Science
Forensic Investigation Center



JOSEPH A. D'AMICO
SUPERINTENDENT

**NEW YORK STATE POLICE
FORENSIC INVESTIGATION CENTER
BUILDING 30
1220 WASHINGTON AVE.
ALBANY, NY 12226-3000**

0.10% SIMULATOR SOLUTION RECORD


Report Date: October 15, 2015
Manufacturer: Guth Laboratories, Inc.
Lot Number: 15210
Expiration Date: 3/21/2016
Analysis Results: 121 mg Ethyl Alcohol / 100 mL of solution

I, Carol S Miller, tested simulator solution lot number 15210 by headspace gas chromatography and have determined that it contains the appropriate concentration of ethyl alcohol. This solution is hereby approved for use.

When this simulator solution is used with a properly operating breath testing instrument, it will provide a value of 0.10% within acceptable limits.

I, Carol S Miller, made the entries contained in this record at the time of each event recorded, or within a reasonable time thereafter.

Performed by:


Digitally signed under
ESRA by Carol S Miller on
2015.10.15 13:52:53 -04'00'

Carol S Miller
Forensic Scientist II
Breath Analysis Operator Permit #43419

Reviewed by:


Digitally signed under
ESRA by Carrie A Kirkton on
2015.10.16 14:04:47 -04'00'

Carrie A Kirkton
Supervisor of Forensic Services
Breath Analysis Technical Supervisor