

CERTIFIED COPY (CPLR RULE 4518)

0.10% SIMULATOR SOLUTION RECORD

Lot Number 25110, dated April 15, 2025

I, Jennifer F Limoges, Associate Director, Forensic Investigation Center, do hereby certify and authenticate, as provided by Rule 4518, subdivisions a and c, Civil Practice Law and Rules, that the document annexed hereto is a true and accurate representation of an electronic record of the New York State Police Forensic Investigation Center which has been delegated to my possession, custody and control by the Superintendent of the New York State Police. This record, once created, is stored in an electronic format that cannot, thereafter, be altered or modified. This record was made in the regular course of business of the New York State Police Forensic Investigation Center, it is the regular course of the Forensic Investigation Center business to make such record, and such record was made at the time of each event recorded in it or within a reasonable time thereafter.

	Digitally signed under ESRA by Jennifer F Limoges on 2025.04.17 07:24:34 -04'00'
--	---

Jennifer F Limoges

Associate Director

Forensic Investigation Center



Division of State Police
Forensic Investigation Center

KATHY HOCHUL
Governor

STEVEN G. JAMES
Superintendent

DR. JILL M. DOOLEY
Director

0.10% SIMULATOR SOLUTION RECORD

Report Date: April 15, 2025
Manufacturer: Guth Laboratories, Inc.
Lot Number: 25110
Expiration Date: 9/25/2025
Analysis Results: 122 mg Ethyl Alcohol / 100 mL of solution

I, Carol S Miller, D-ABFT-FA, tested simulator solution lot number 25110 by headspace gas chromatography and have determined that it contains the appropriate concentration of ethyl alcohol. This solution is hereby approved for use.

When this simulator solution is used with a properly operating breath testing instrument, it will provide a value of 0.10% within acceptable limits.

I, Carol S Miller, D-ABFT-FA, made the entries contained in this record at the time of each event recorded, or within a reasonable time thereafter.

Performed by:

 Digitally signed under
ESRA by Carol S Miller on
2025.04.15 10:13:17 -04'00'

Carol S Miller, D-ABFT-FA

Forensic Scientist III

Breath Analysis Technical Supervisor/Operator #43419

Reviewed by:

 Digitally signed under
ESRA by Carrie A Kirkton
on
2025.04.15 13:40:48 -04'00'

Carrie A Kirkton, D-ABFT-FA

Supervisor of Forensic Services

Breath Analysis Technical Supervisor