

SUPERVISOR TEST

 GLENS FALLS CITY
 POLICE DEPT
 DATAMASTER dmt: 117106

Date: 06/08/2026

OPERATOR NAME:
 JOHN T MULCAHY
 Permit #: 54048
 Operator Agency/Dept: GLENS FALLS CITY PD
 Reference Standard #: 53283

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:48
INTERNAL STANDARD	VERIFIED	04:48
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:49
BLANK TEST	0.00	04:50
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:50
BLANK TEST	0.00	04:51
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:51
BLANK TEST	0.00	04:52

