



Ryan S. Pedone

Chief of Police

Telephone: (518) 761-3840

Facsimile: (518) 798-4345

42 Ridge Street, Glens Falls, New York 12801

To: Warren County District Attorney's Office
From: Chief Ryan Pedone
Date: 8/20/2026
Subject: Glens Falls Police Datamaster DMT

On July 15th 2026 we had a Fitler Wheel Error on our Datamaster DMT while conducting a weekly test. The Datamaster DMT was taken out of service and taken to DCJS – Highway Safety Technology Unit for repair. On August 20th 2026 the Datamaster DMT was placed back in service, and a weekly test was ran.

A handwritten signature in black ink, appearing to be "R. Pedone", written in a cursive style.

Chief of Police
Ryan S. Pedone