

SOLUTION CHANGE

 GLENS FALLS CITY
 POLICE DEPT
 DATAMASTER dmt: 117106

Date: 04/01/2026

OPERATOR NAME:
 HUNTER L BARTON
 Permit #: 49353
 Operator Agency/Dept: GLENS FALLS CITY PD
 Reference Standard #: 26050

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:02
INTERNAL STANDARD	VERIFIED	07:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	07:03
BLANK TEST	0.00	07:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	07:04
BLANK TEST	0.00	07:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	07:05
BLANK TEST	0.00	07:06

