

SUPERVISOR TEST

 GLENS FALLS CITY
 POLICE DEPT
 DATAMASTER dmt: 117106

Date: 04/06/2026

OPERATOR NAME:
 RICHARD R COLLODI
 Permit #: 55571
 Operator Agency/Dept: GLENS FALLS CITY PD
 Reference Standard #: 26050

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:31
INTERNAL STANDARD	VERIFIED	02:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:31
BLANK TEST	0.00	02:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:32
BLANK TEST	0.00	02:33
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:34
BLANK TEST	0.00	02:34

