

SUPERVISOR TEST

 GLENS FALLS CITY
 POLICE DEPT
 DATAMASTER dmt: 117106

Date: 04/20/2026

OPERATOR NAME:
 REESE P WHEATON
 Permit #: 53611
 Operator Agency/Dept: GLENS FALLS CITY PD
 Reference Standard #: 26050

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:10
INTERNAL STANDARD	VERIFIED	03:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:11
BLANK TEST	0.00	03:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:12
BLANK TEST	0.00	03:13
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:13
BLANK TEST	0.00	03:14

