

SUPERVISOR TEST

 GLENS FALLS CITY
 POLICE DEPT
 DATAMASTER dmt: 117106

Date: 05/25/2026

OPERATOR NAME:
 RICHARD R COLLODI
 Permit #: 55571
 Operator Agency/Dept: GLENS FALLS CITY PD
 Reference Standard #: 26050

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:29
INTERNAL STANDARD	VERIFIED	01:29
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:30
BLANK TEST	0.00	01:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:31
BLANK TEST	0.00	01:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:32
BLANK TEST	0.00	01:33

