

SUPERVISOR TEST

 GLENS FALLS CITY
 POLICE DEPT
 DATAMASTER dmt: 117106

Date: 06/15/2026

OPERATOR NAME:
 JOHN T MULCAHY
 Permit #: 54048
 Operator Agency/Dept: GLENS FALLS CITY PD
 Reference Standard #: 53283

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:31
INTERNAL STANDARD	VERIFIED	05:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:31
BLANK TEST	0.00	05:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:32
BLANK TEST	0.00	05:33
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:34
BLANK TEST	0.00	05:34

