

SOLUTION CHANGE

 GLENS FALLS CITY
 POLICE DEPT
 DATAMASTER dmt: 117106

Date: 07/01/2026

OPERATOR NAME:
 DANELLE M BELL
 Permit #: 53551
 Operator Agency/Dept: GLENS FALLS CITY PD
 Reference Standard #: 26050

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:00
INTERNAL STANDARD	VERIFIED	01:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:00
BLANK TEST	0.00	01:01
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:01
BLANK TEST	0.00	01:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:03
BLANK TEST	0.00	01:04

