

SUPERVISOR TEST

 GLENS FALLS CITY
 POLICE DEPT
 DATAMASTER dmt: 117106

Date: 07/06/2026

OPERATOR NAME:
 RICHARD R COLLODI
 Permit #: 55571
 Operator Agency/Dept: GLENS FALLS CITY PD
 Reference Standard #: 26120

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:02
INTERNAL STANDARD	VERIFIED	03:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:03
BLANK TEST	0.00	03:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:04
BLANK TEST	0.00	03:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:05
BLANK TEST	0.00	03:06

