

SUPERVISOR TEST

 GLENS FALLS CITY
 POLICE DEPT
 DATAMASTER dmt: 117106

Date: 07/13/2026

OPERATOR NAME:
 HUNTER L BARTON
 Permit #: 49353
 Operator Agency/Dept: GLENS FALLS CITY PD
 Reference Standard #: 26120

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:39
INTERNAL STANDARD	VERIFIED	04:39
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:40
BLANK TEST	0.00	04:40
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:41
BLANK TEST	0.00	04:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:42
BLANK TEST	0.00	04:43

