

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 08/20/2026

OPERATOR NAME:
JOHN T MULCAHY
Permit #: 54048
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 26190

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:32
INTERNAL STANDARD	VERIFIED	00:33
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:33
BLANK TEST	0.00	00:34
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:34
BLANK TEST	0.00	00:35
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:35
BLANK TEST	0.00	00:36

