

SOLUTION CHANGE

 GLENS FALLS CITY
 POLICE DEPT
 DATAMASTER dmt: 117106

Date: 09/01/2026

OPERATOR NAME:
 JOHN T MULCAHY
 Permit #: 54048
 Operator Agency/Dept: GLENS FALLS CITY PD
 Reference Standard #: 26190

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:29
INTERNAL STANDARD	VERIFIED	23:29
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:29
BLANK TEST	0.00	23:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:30
BLANK TEST	0.00	23:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:32
BLANK TEST	0.00	23:32

