

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 09/16/2026

OPERATOR NAME:
HUNTER L BARTON
Permit #: 49353
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 26190

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:12
INTERNAL STANDARD	VERIFIED	07:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	07:12
BLANK TEST	0.00	07:13
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	07:14
BLANK TEST	0.00	07:15
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	07:15
BLANK TEST	0.00	07:16

