

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 09/21/2026

OPERATOR NAME:
REESE P WHEATON
Permit #: 53611
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 26190

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:18
INTERNAL STANDARD	VERIFIED	01:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:19
BLANK TEST	0.00	01:20
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:20
BLANK TEST	0.00	01:21
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:21
BLANK TEST	0.00	01:22

