

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 09/07/2026

OPERATOR NAME:
HUNTER L BARTON
Permit #: 49353
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 26190

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:46
INTERNAL STANDARD	VERIFIED	06:46
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	06:46
BLANK TEST	0.00	06:47
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	06:48
BLANK TEST	0.00	06:48
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	06:49
BLANK TEST	0.00	06:50

