

SOLUTION CHANGE




 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 144506

Date: 11/04/2013

Operator Name: RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 13200

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:31
INTERNAL STANDARD	VERIFIED	17:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:32
BLANK TEST	0.00	17:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:33
BLANK TEST	0.00	17:33
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	17:34
BLANK TEST	0.00	17:35



SG *RJ Bartlett* *R Bartlett* WCSO #102 - 11/4/13