

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 10/14/2020

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:01
INTERNAL STANDARD	VERIFIED	14:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:01
BLANK TEST	0.00	14:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:03
BLANK TEST	0.00	14:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:04
BLANK TEST	0.00	14:05



Charles B. Geisler #241