

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 03/06/2026

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:49
INTERNAL STANDARD	VERIFIED	17:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:49
BLANK TEST	0.00	17:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:50
BLANK TEST	0.00	17:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:52
BLANK TEST	0.00	17:52

PD Steven T Bunio #189

