

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 03/15/2026

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:20
INTERNAL STANDARD	VERIFIED	03:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:21
BLANK TEST	0.00	03:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:22
BLANK TEST	0.00	03:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	03:23
BLANK TEST	0.00	03:24

[Handwritten signature]
#820

