

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 03/27/2026

OPERATOR NAME:
TREVOR C HOPECK
Permit #: 49898
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:37
INTERNAL STANDARD	VERIFIED	22:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:37
BLANK TEST	0.00	22:38
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	22:38
BLANK TEST	0.00	22:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:39
BLANK TEST	0.00	22:40

T. Hopec #101

