

# SOLUTION CHANGE

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 04/03/2026

OPERATOR NAME:  
ANTHONY E CARDENAS  
Permit #: 50642  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 26050

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 00:30 |
| INTERNAL STANDARD       | VERIFIED | 00:30 |
| REFERENCE STANDARD TEMP | 34.00c   |       |
| REFERENCE STANDARD      | 0.09     | 00:30 |
| BLANK TEST              | 0.00     | 00:31 |
| REFERENCE STANDARD TEMP | 34.01c   |       |
| REFERENCE STANDARD      | 0.10     | 00:31 |
| BLANK TEST              | 0.00     | 00:32 |
| REFERENCE STANDARD TEMP | 34.00c   |       |
| REFERENCE STANDARD      | 0.10     | 00:32 |
| BLANK TEST              | 0.00     | 00:33 |

*184*

