

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 04/10/2026

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 26050

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:50
INTERNAL STANDARD	VERIFIED	21:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	21:51
BLANK TEST	0.00	21:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	21:52
BLANK TEST	0.00	21:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	21:53
BLANK TEST	0.00	21:54

#820

