

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 06/01/2026

OPERATOR NAME:
J T COON
Permit #: 36043
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 26120

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:33
INTERNAL STANDARD	VERIFIED	22:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:33
BLANK TEST	0.00	22:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:34
BLANK TEST	0.00	22:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:36
BLANK TEST	0.00	22:36



Sgt J.T. Coon *[Signature]* 341