

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 03/06/2026

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:18
INTERNAL STANDARD	VERIFIED	07:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:18
BLANK TEST	0.00	07:19
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:20
BLANK TEST	0.00	07:20
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:21
BLANK TEST	0.00	07:22

[Handwritten signature]
44983

