

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 03/15/2026

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:00
INTERNAL STANDARD	VERIFIED	01:00
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:00
BLANK TEST	0.00	01:01
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:01
BLANK TEST	0.00	01:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:02
BLANK TEST	0.00	01:03

Handwritten signature
#820

