

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 05/16/2026

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 26050

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:40
INTERNAL STANDARD	VERIFIED	09:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:41
BLANK TEST	0.00	09:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	09:42
BLANK TEST	0.00	09:43
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	09:43
BLANK TEST	0.00	09:44

[Handwritten signature]

