

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 05/30/2026

OPERATOR NAME:
JOHN D BATEHOLTS
Permit #: 43007
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 26050

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:52
INTERNAL STANDARD	VERIFIED	13:52
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:53
BLANK TEST	0.00	13:54
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:54
BLANK TEST	0.00	13:55
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:55
BLANK TEST	0.00	13:56



Handwritten signature and date: 07/11