

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 06/01/2026

OPERATOR NAME:
J T COON
Permit #: 36043
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 26120

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:33
INTERNAL STANDARD	VERIFIED	20:33
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:33
BLANK TEST	0.00	20:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:35
BLANK TEST	0.00	20:35
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:36
BLANK TEST	0.00	20:37



Sgt J.T. Coon *[Signature]* 341