

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 03/15/2026

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:16
INTERNAL STANDARD	VERIFIED	04:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	04:16
BLANK TEST	0.00	04:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	04:17
BLANK TEST	0.00	04:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	04:19
BLANK TEST	0.00	04:19

#870

