

SUPERVISOR TEST




WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 05/30/2026

OPERATOR NAME:
JOHN D BATEHOLTS
Permit #: 43007
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 26050

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:21
INTERNAL STANDARD	VERIFIED	14:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:21
BLANK TEST	0.00	14:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:22
BLANK TEST	0.00	14:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:24
BLANK TEST	0.00	14:25



[Handwritten signature] #714