

SOLUTION CHANGE




WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 06/01/2026

OPERATOR NAME:
J T COON
Permit #: 36043
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 26120

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:02
INTERNAL STANDARD	VERIFIED	23:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:02
BLANK TEST	0.00	23:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:04
BLANK TEST	0.00	23:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:05
BLANK TEST	0.00	23:06



SGT J.T. Coon *[Signature]* 341